

04 October 2011

Ref: SM/AM/JL/2011

**For the attention of Jacqy Lim  
ASIA MARINE SUPPLIES PTE LTD**

To Whom It May Concern,

**RE: The analysis of "FLEXSEALS UNIGRAPH 500" for the presence of asbestos under polarised light microscopy (PLM) with dispersion staining.**

Upon receipt of the above sample, Acorn Analytical Services Limited, which is an UKAS Accredited Testing Laboratory, undertakes analysis for the presence of asbestos using Polarised Light Microscopy (PLM) with Dispersion Staining.

*The results from the bulk analysis confirmed that **Flexseals Unigraph 500** is an ASBESTOS FREE product. The accompanying bulk certificate (**PRN3716**) follows this letter.*

Should you need further assistance, please contact your vendor accordingly.

Yours sincerely



Acorn Analytical Services  
**Stephen McGlone**  
Operations Director

## Certificate Of Bulk Analysis For Asbestiform Material


Client: <input type="text" value="Asia Marine Supplies Pte Ltd"/>	PRN No: <input type="text" value="PRN3716"/>
Contact: <input type="text" value="Jacqy Lim"/>	Site: <input type="text" value="FLEXSEALS UNIGRAPH 500"/>
Client Address: <input type="text" value="23 Pioneer Sector 1&lt;br/&gt;Singapore 628431"/>	
Contact No: <input type="text" value="62664388"/>	
Sample Taken by: <input type="text" value="Not Specified"/>	OF: <input type="text" value="Not Specified"/>
Date taken: <input type="text"/>	Received Date: <input type="text" value="03/10/2011"/> Analysis Date: <input type="text" value="03/10/2011"/>

CD61 Issue 1 Revision 4 09/11/10

The samples were analysed using polarised light microscopy with dispersion staining in accordance with Acorn Analytical Services Ltd documented in-house procedures based upon HSE document "The Analyst Guide". Where Acorn Analytical Services Ltd did not take sample(s), the results given are based upon information supplied by those taking the sample(s). In this instance, Acorn Analytical Services Ltd guarantees the accuracy of the sample analysis only.

This test report should not be reproduced, except in full, without written permission from Acorn Analytical Services Ltd. Opinions and interpretations raised on this certificate are outside the scope of UKAS accreditation.

## Analysis Details

Samples Analysis By: <input type="text" value="Jonathan Parker"/>	Analyst Signature: 
Position: <input type="text" value="Environmental Consultant"/>	
Issued by: <input type="text" value="Jonathan Parker"/>	Issued by Signature: 
Issued by position: <input type="text" value="Environmental Consultant"/>	

## Analysis Results

Sample Ref	Material Type	Sample Details	Result
BR96667	Gasket	FLEXSEALS UNIGRAPH 500	Non detected